



Factor II, Incorporated

Inventing and Innovating...

Information: 1.928.537.8387 www.factor2.com

ATTESTATION

Factor II, Inc. RTV Medical Adhesive A-100

Factor II, Inc. RTV Medical Adhesive A-100 has undergone cytotoxicity testing per NAmSA report 09C_31445_01 – 2009. Tested material successfully passed the test conducted and was found to be non-cytotoxic.

Bill McFall
Quality Assurance Manager
Factor II, Inc.